



**“NOVEL MINIATURIZED ANTENNAS AND ARRAYS FOR
IMPLANTABLE, INGESTIBLE AND BODY-WORN
APPLICATIONS”**

*A thesis submitted
in fulfilment of the requirements for the degree of
Doctor of Philosophy (PhD)*

by

Farhad Gozasht

University of Technology Sydney

June 2018

Certificate of Original Authorship

I certify that the work in this thesis has not previously been submitted for a degree nor has it been submitted as part of requirements for a degree except as fully acknowledged within the text.

I also certify that the thesis has been written by me. Any help that I have received in my research work and the preparation of the thesis itself has been acknowledged. In addition, I certify that all information sources and literature used are indicated in the thesis.

This research is supported by an Australian Government Research Training Program.

Production Note:

Signature of Student: Signature removed prior to publication.

Date: 12.June.2018

ACKNOWLEDGEMENTS

There are many people, whom I am indebted to, without whose assistance, this thesis would not have been possible. I would like to take this opportunity to express my outmost gratitude to them.

I consider it my privilege to work with A/Prof. Ananda M. Sanagavarapu. His motivation, supervision and commitment to his beliefs in strengthening my fundamentals have been instrumental in my growth.

I would like to explicitly thank Mr. Russell Nicholson for his efforts in providing me with the measurement facility and safe experimental environment without which, this thesis would not be in its present state. I wish to express my heartfelt gratitude to Mr. Raymond Clout for his support and providing the SAR measurement tools, that added value to my research work.

Furthermore, the CHT group needs a special mention here, for standing by me and making my life here eventful.

Last, but never the least, I am very grateful to my family, who have always worked hard to help me reach my goals without much discomfort.

Above all, I thank the Almighty for showering me with His Grace & Blessings.

Farhad Gozasht

Contents

CERTIFICATE OF ORIGINAL AUTHORSHIP	i
ACKNOWLEDGEMENTS.....	ii
LIST OF FIGURES	viii
LIST OF TABLES.....	xv
LIST OF ACRONYMS USED	xvii
LIST OF SYMBOLS USED	xxi
ABSTRACT.....	xxii

CHAPTER – 01

INTRODUCTION

1.1 Principal characteristic of wireless body-worn and implantable systems for EM energy deposition and data biotelemetry.....	4
1.1.1 Frequency bands and spectrum allocation	5
1.1.2 Human body.....	7
1.1.3 Array of antennas.....	7
1.1.4 Feeding network and coupling bolus	8
1.1.5 SAR and temperature rise	8
1.1.6 Base station.....	9
1.1.7 Implantable antennas	9
1.1.8 Channel propagation.....	9
1.1.9 Insulations.....	10
1.1.10 Bio actuators and Bio sensors.....	10
1.1.11 Electronics and power supply	10
1.1.12 Experiments and characterization.....	11
1.2 Thesis outlook.....	11
1.2.1 Thesis outline and original contributions.....	12

CHAPTER – 02

A REVIEW OF EXISTING LITERATURE

2.1 Background.....	17
2.2 Implantable and wearable antennas	19
2.2.1 Implantable antennas: applications and miniaturization.....	20

CHAPTER – 03

DUAL SLOT PIFAs FOR BODY IMPLANTABLE DEVICES

3.1 Introduction.....	42
3.2 Background review and motivation.....	43
3.3 Design of novel PIFA implantable antennas	49
3.3.1 Initial design	49
3.3.2 Parametric study	55
3.3.3 Antenna performance.....	62
3.4 DS-PIFA measurements	66
3.5 Specific Absorption Ratio (SAR)	73
3.6 Biotelemetry and wireless communication link.....	77
3.6.1 Statistical Path Loss Model	78
3.6.2 Pathloss measurement.....	79
3.6.3 Link budget.....	81
3.7 DS-PIFAs performance with different slots arrangement, Case I: (Slot position on two neighbouring sides).....	83
3.8 Discussion.....	89

CHAPTER – 04

TRANSMITTER AND RECEIVER ANTENNA FOR INGESTIBLE WIRELESS CAPSULE ENDOSCOPY SYSTEMES

4.1 Introduction.....	91
4.2 Background review and motivation.....	93

4.3 Modelling of human GI tract environments	97
4.4 Compact Tx antenna to fit into capsule compartment.....	97
4.5 Wideband folded co-planar loop antenna	100
4.5.1 Technique for improved antenna design.....	102
4.5.2 Design of printed folded co-planar loop antenna	104
4.6 Measurements for the proposed antenna	109
4.6.1 SAR Measurements	111
4.7 Body-worn antenna for base station array	114
4.7.1 Numerical phantoms for simulations.....	115
4.7.2 Design procedure of Compact CP-PIFA	116
4.7.3 Measurements on Compact CP-PIFA.....	120
4.8 Receiver array for wireless endoscopy systems	122
4.8.1 Numerical phantoms for simulation	123
4.8.2 Antenna array design	124
4.8.3 Measurements	126
4.9 Development of path loss model for heterogeneous human body based on received signal data.....	128
4.9.1 Setup for homogeneous body tissue: Insulated dipoles.....	128
4.9.2 Setup for inhomogeneous body model: half-wavelength dipoles	134
4.9.3 Pathloss calculation using compact CP-PIFA as Rx antennas....	137
4.9.4 Experimental validation.....	139
4.10 Discussion	142

CHAPTER – 05

MINIATURIZED ANTENNAS FOR HEAD IMPLANTABLE DEVICES

5.1 Introduction.....	144
5.2 Background review and motivation.....	145

5.3 Single Slot PIFA design and performance.....	151
5.3.1 Design procedure	152
5.3.2 Calculation environment.....	156
5.3.3 Parametric study	159
5.3.4 Antenna performance.....	163
5.4 SS-PIFA measurements	165
5.5 Specific Absorption Ratio (SAR)	168
5.6 Folded oval loop	173
5.7 Multi band Single Slot PIFA	178
5.7.1 Design of multi band Single Sslot PIFA.....	179
5.7.2 Antenna performance and numerical phantom.....	182
5.7.3 Experimental results	184
5.8 Discussion.....	188

CHAPTER – 06

NEAR FIELD FOCUSED BODY WORN ARRAY FOR NON-INVASIVE MICROWAVE HYPERTHEMIA

6.1 Introduction.....	190
6.2 Background review and motivation	191
6.3 Antenna design	196
6.3.1 Calculation environment.....	197
6.3.2 Specific Absorption Rate (SAR) and Bio-heat transfer equation.....	200
6.3.3 Array configuration.....	201
6.3.4 Technics to optimize focusing	209
6.3.5 Array element design	214
6.4 Calculation results.....	218
6.5 Antenna array with limited number of elements for NI-LMH	221

6.5.1	Square arrangement	222
6.5.2	Circular arrangement	223
6.5.3	Cross arrangement	224
6.5.4	Hexagonal arrangement	226
6.5.5	Outer square & inner circular arrangement	227
6.5.6	Outer circular and inner square arrangement.....	229
6.6	Experimental setup and temperature measurement	232
6.7	Discussion	241

CHAPTER – 07

CONCLUSION AND FUTURE WORK

7.1	Conclusion	244
7.2	Future work.....	259

PAPERS PUBLISHED SO FAR.....	262
------------------------------	-----

REFERENCES	264
------------------	-----

LIST OF FIGURES

Figure 1-1. Example of monitoring home healthcare system using wireless implantable devices.....	3
Figure 1-2. Different aspects from antenna engineer point of view for microwave cancer treatment system.	4
Figure 1-3. Different aspects from antenna engineer point of view for wireless implantable system for data biotelemetry.	5
Figure 3-1: Top views of proposed DS-PIFA geometry with different slot position (a), (b) and (c).....	51
Figure 3-2: Flowchart for initial antenna design methodology.....	51
Figure 3-3: Simulation set up in numerical mimicking phantoms, a) 3D complex human body model, b) Side view of layered Block and Cylindrical phantoms.	52
Figure 3-4: Simulation set up in numerical mimicking phantoms, a) Multi-Layered Block phantom, and b) Cylindrical phantom.....	54
Figure 3-5: The method followed for DS-PIFAs size reduction.....	57
Figure 3-6: The surface current distribution at 2.45 GHz (top view), a) first prototype (b) perturbed surface current (before adjusting feeding and pin position), c) miniaturized prototype, d) miniaturized size with smaller ground plane.....	58
Figure 3-7: Ground plane size variation versus antenna bandwidth.....	59
Figure 3-8: Dual-slot PIFA: The effects of reducing the slot width.	61
Figure 3-9: The electric near field distribution (top view), a) outside the numerical phantom, b) inside the skin layer, c) inside the muscle layer and d) side view.....	63
Figure 3-10: Simulation environment for possible air gap layer and interstitial fluid when implant DS-PIFAs in human body.....	66
Figure 3-11: comparison of relative permittivity and conductivity of the beef phantom with published data in skin and muscle.	66

Figure 3-12: Reflection coefficient of Dual-slot PIFA: a) Calculated using Complex human body model, block Layered and cylindrical Layered phantoms, b) Measured vs. Calculated using Complex human body model.....	67
Figure 3-13: Measurement set up and Dual-Slot PIFAs prototype on FR4 and Rogers 3210 substrate for different frequency bands.	68
Figure 3-14: Reflection coefficient of Dual-slot PIFA covering ISM at 2400MHz using Rogers substrate: Calculated using block layered and homogeneous human tissue numerical mimicking phantoms versus Measured inside experimental phantom.	69
Figure 3-15: Reflection coefficient of Dual-slot PIFA covering WMTS at 1400MHz using FR4 substrate: Calculated using block layered and homogeneous human tissue numerical mimicking phantoms versus Measured inside experimental phantom.....	70
Figure 3-16: Reflection coefficient of Dual-slot PIFA covering WMTS at 1400MHz using Rogers substrate: Calculated using block layered and homogeneous human tissue numerical mimicking phantoms versus Measured inside experimental phantom.	70
Figure 3-17: Reflection coefficient of Dual-slot PIFA covering Zigbee at 868MHz using FR4 substrate: Calculated using block layered and homogeneous human tissue numerical mimicking phantoms versus Measured inside experimental phantom.....	71
Figure 3-18: Reflection coefficient of Dual-slot PIFA covering Zigbee at 868MHz using Rogers substrate: Calculated using block layered and homogeneous human tissue numerical mimicking phantoms versus Measured inside experimental phantom.	71
Figure 3-19: Reflection coefficient of Dual-slot PIFA covering MICS band at 400MHz using FR4 substrate: Calculated using block layered and homogeneous human tissue numerical mimicking phantoms versus Measured inside experimental phantom.....	72
Figure 3-20: Reflection coefficient of Dual-slot PIFA covering MICS band at 400MHz using Rogers substrate: Calculated using block layered and	

homogeneous human tissue numerical mimicking phantoms versus Measured inside experimental phantom.	72
Figure 3-21: a) SAR measurement diagram using optical fibre thermal sensors. b) SAR measurement setup for Dual-slot PIFA on FR4 substrate.	74
Figure 3-22: Measured 1g-SAR: Miniaturized Dual-Slot PIFA at 2.4 GHz.	75
Figure 3-23: Simulation environment of the wireless power link.....	79
Figure 3-24: Calculated and Measured pathloss variation for Rx and Tx separation.	80
Figure 3-25: Pathloss measurement inside the UTS FEIT Building (Building 11)...80	
Figure 3-26: DS-PIFA slot position, Case I, on two neighbouring sides, left and bottom.	83
Figure 3-27: Different slots' location for DS-PIFA (Case I, Slot position on two neighbouring sides).....	84
Figure 3-28: DS-PIFA, Case I: Horizontal slot length variation.	85
Figure 3-29: DS-PIFA, Case I: Vertical slot length variation.....	86
Figure 3-30: DS-PIFA, Case I: Horizontal and vertical slots length variation.	87
Figure 3-31: DS-PIFA, Case I: Extended vertical slot length variation.	89
Figure 4-1. Simulation setup: Block-layered arrangement.	98
Figure 4-2. HFSS simulation setup: Complex human torso with GI tract.	98
Figure 4-3. Implanted sensor conception: Folded co-planar loop antenna inside the capsule for endoscopy systems. Maximum dimensions are indicated. Placements of battery and electronics are indicated to estimate the overall volume occupancy.	99
Figure 4-4. Wide band Tx loop antenna geometry: Unfolded version.	102
Figure 4-5. Wide band Tx antenna geometry: folded co-planar loop.....	104
Figure 4-6. Wide band Tx antenna geometry: Printed folded co-planar loop antenna.	106
Figure 4-7. VSWR comparison of folded and modified loop antennas.....	107
Figure 4-8. Surface current distribution for the folded co-planar loop antenna.....	108
Figure 4-9. Transmitter antenna on FR4: Top view and bottom view.....	109

Figure 4-10. Measurement setup for Transmitter antenna on FR4.....	110
Figure 4-11. The comparison of relative permittivity and conductivity of the experimental phantom, skin and small intestine.....	110
Figure 4-12. Reflection coefficient comparison: calculated versus measured.....	111
Figure 4-13. Schematic of optical fibre sensors arrangement for 1g-SAR measurement.	112
Figure 4-14. 1g-SAR measurement setup using optical fibre sensors arrangement.	112
Figure 4-15. SAR calculation for transmitter antenna inside the small intestine.....	113
Figure 4-16. Geometry of proposed receiver antenna.....	115
Figure 4-17. Simulation setup: CP-PIFA placed close to the Complex human torso.	116
Figure 4-18. Receiver element surface current distribution at 403 MHz.....	118
Figure 4-19. Receiver antenna axial ratio (A.R).....	118
Figure 4-20. Receiver antenna: prototype and measurement setup.	121
Figure 4-21. The reflection coefficient and S parameter comparison for CP-PIFA receiver antenna: Measurement versus simulation.	122
Figure 4-22. Simulation setup to calculate antenna performance: The HFSS complex human body numerical model and simplified cylindrical homogeneous human tissue phantom.	124
Figure 4-23. Receiver array element prototypes and measurement setup.	127
Figure 4-24. Power received at six receiver ports.....	127
Figure 4-25. Side View (left) and Top View (right) of the Simulation Setup using dipoles as receivers and printed loop antenna as transmitter.....	130
Figure 4-26. Scatter Plot of Path Loss Data.....	133
Figure 4-27. Setup for calculating scattering parameters data using half-wavelength dipoles as receiver. a) Complex human body phantom, b) The position of Rx antennas (in red) and Tx antenna (in blue) around the body model.	135
Figure 4-28. Scatter Plot of Path Loss Data.....	136

Figure 4-29. Setup for calculating scattering parameters using CP-PIFA as array antenna receiving element. Rx arrays are placed in three rings consisting of six CP-PIFAs each (total of 18 antennas).....	137
Figure 4-30. Scatter Plot of Path Loss Data.....	138
Figure 4-31. Setup for calculating scattering parameters for the different positions of transmitter and receiver antennas. Rx arrays are placed in three rings consisting of six CP-PIFAs each.	139
Figure 4-32. Setup for collecting the $ S_{12} _{[dB]}$ measured data for the proposed Tx and Rx antennas. Total of six receiver antennas are used.	140
Figure 4-33. Path loss data for Tx antenna position-1 inside the meat phantom. The lowest path loss recorded is around 15 dB while the highest is around 45dB.....	140
Figure 4-34. The side view of the measurement setup for the $ S_{12} _{[dB]}$ data collection using the array of six compact CP-PIFAs for three different positioning of Tx antenna.	141
Figure 5-1: Top view of SS-PIFA geometry.....	153
Figure 5-2: The initial antenna design flowchart.	153
Figure 5-3: Simulation set up in numerical mimicking phantoms, a) 3D complex human head model, b) Side view of layered Block and cylindrical phantoms.	154
Figure 5-4. Single Slot PIFA structure.....	155
Figure 5-5: Simulation set up in numerical mimicking phantoms, a) 3D complex human head model, b) Side view of Multi-Layered Block, Multi-Layered Spherical and cylindrical phantoms.	157
Figure 5-6: The SS-PIFAs size reduction flowchart.	160
Figure 5-7: Single slot PIFA: The effects of reducing the slot width.	161
Figure 5-8: Reflection coefficient of Single slot PIFA covering ISM band when immersed inside the multilayer spherical numerical phantom and the complex human body head phantom.	163
Figure 5-9: Measurement set up and Single Slot PIFAs prototype, using etching technique, on FR4 substrate for different frequency bands.	166

Figure 5-10: Single Slot PIFAs prototype on FR4 substrate for different frequency bands of interest.	167
Figure 5-11: comparison of relative permittivity and conductivity of the beef phantom with published data in skin and muscle.	168
Figure 5-12: Reflection coefficient of Single slot PIFA: a) Calculated using Complex human body model, block Layered and cylindrical Layered phantoms, b) Measured vs. Calculated using Complex human body model.....	169
Figure 5-13: a) SAR measurement diagram using optical fibre thermal sensors. b) SAR measurement setup for Single slot PIFA on FR4 substrate.	170
Figure 5-14: Measured 1g-SAR: Miniaturized Single Slot PIFA at 2450 MHz.....	171
Figure 5-15. Wide band oval loop antenna geometry: Unfolded loop.....	174
Figure 5-16. Wide band folded oval loop antenna geometry.....	175
Figure 5-17. Calculated reflection coefficient of compact folded oval loop antenna.	176
Figure 5-18. Surface current distribution for the folded co-planar loop antenna.....	177
Figure 5-19. Top view of SS-PIFA geometry.....	181
Figure 5-20. Triple band Single Slot PIFA structure.	181
Figure 5-21. Complex human body model.	183
Figure 5-22. Surface current distribution of the proposed antenna at (a) MICS band and (b) WMTS band.	184
Figure 5-23. Triple band Single Slot PIFA prototype and Measurement set up.....	185
Figure 5-24. Reflection coefficient: Measured in pork skin vs. Simulation in layered block and cylindrical tissue phantoms.	187
Figure 5-25. Reflection coefficient: Measured in pork skin vs. simulation using inhomogeneous tissue phantom.	187
Figure 6-1: Simulation set up in numerical mimicking phantom: 3D complex human body model.	199
Figure 6-2. Calculation set up for the proposed array inside the bolus close by the multi-layer numerical phantom (side view).....	199

Figure 6-3: a) The presented linear array immersed inside the coupling bolus. b) The applicator package close by the multi-layer block human tissue mimicking phantom.	203
Figure 6-4. E-field distribution of linear array at depth of 5 cm.	204
Figure 6-5. Circular array configuration.	205
Figure 6-6. E-field distribution of circular array at depth of 5 cm.	205
Figure 6-7. A three-dimensional view of the planar square array geometry [108]. ..	208
Figure 6-8. E-field distribution of array at depth of 5 cm.	208
Figure 6-9. Configuration of the proposed array of antenna array inside the bolus (top view).	208
Figure 6-10. Antenna array and receiving antenna separated by distance D.	210
Figure 6-11. Geometry of Dual Slot PIFA.	215
Figure 6-12. The DS-PIFAs size reduction flowchart.	216
Figure 6-13. Printed Dual Slot PIFA on FR4 substrate and measurement setup.	216
Figure 6-14. Measured vs calculated: Homogeneous numerical phantom.	217
Figure 6-15. Temperature distribution of 5×5 planar array.	219
Figure 6-16. Normalized SAR distribution at depth of 3.5 cm.	219
Figure 6-17. E-field distribution of array at different target depth along z axis: a) 20mm, b) 30mm, c) 40mm, d) 50mm from the body surface.	220
Figure 6-18. Array of antenna (feeding cables included) inside the applicator for NI-LMH.	221
Figure 6-19. Square array arrangement: a) without parasitic antenna, b) with parasitic antenna.	222
Figure 6-20. E-field distribution of square array arrangement.	223
Figure 6-21. Circular array arrangement: a) without parasitic antenna, b) with parasitic antenna.	224
Figure 6-22. E-field distribution of circular array arrangement.	224
Figure 6-23. Cross array arrangement: a) without parasitic antenna, b) with parasitic antenna.	225
Figure 6-24. E-field distribution of cross array arrangement.	225

Figure 6-25. Hexagonal array arrangement: a) without parasitic antenna, b) with parasitic antenna.	227
Figure 6-26. E-field distribution of hexagonal array arrangement.	227
Figure 6-27. Outer Square & Inner Circular array arrangement: a) without parasitic antenna, b) with parasitic antenna.....	228
Figure 6-28. E-field distribution of Outer Square & Inner Circular array arrangement.	228
Figure 6-29. Inner Square & Outer Circular array arrangement: a) without parasitic antenna, b) with parasitic antenna.....	230
Figure 6-30. Temperature distribution of the nine elements planar array.....	231
Figure 6-31. E-field distribution of Outer Circular & Inner Square array arrangement at depth of 5 cm.	231
Figure 6-32. Temperature raise measurement block diagram.....	233
Figure 6-33. Comparison of relative permittivity and conductivity of the beef phantom with published data in skin and muscle.	234
Figure 6-34. Temperature measurement points at depth of 3 cm from the phantom surface on x-y plane: a) desired hot spot in red, b) surrounding tissue and boundary points in green.....	234
Figure 6-35. Printed NFF antenna array prototype: a) Top view and b) Bottom view.	235
Figure 6-36. Temperature raise measurement setup.	235
Figure 6-37. Reflection coefficient of the port 1 and 2 of the proposed NFF array: Calculated using Complex human body model vs Measured using experimental phantom.....	236
Figure 6-38. Power received at six receiver ports.....	236
Figure 6-39. Printed NFF antenna array prototype: a) Top view and b) Bottom view.	238
Figure 6-40. Printed NFF antenna array prototype: a) Top view and b) Bottom view.	239

LIST OF TABLES

Table 1-1 Frequency spectrum allocation for biomedical applications	7
Table 2-1 Performance comparisons for the state of MICS-Band implantable antennas	26
Table 2-2 Performance comparisons for the state of art ISM-Band implantable antennas.....	28
Table 2-3 Size and performance comparisons for the state of art capsule Tx antenna	31
Table 2-4 Size and performance comparisons for the state of art capsule Rx antenna	34
Table 3-1 DS-PIFAs size and characteristics over different medical frequency bands	55
Table 3-2 Size comparison of proposed Dual Slot PIFA.....	60
Table 3-3 Miniaturized DS-PIFAs size and characteristic over different medical frequency bands	62
Table 3-4 Losses in DS-PIFA and numerical phantom at 2.4 GHz (in watt)	645
Table 3-5 Variations of tissue environment during implantation.....	65
Table 3-6 Variations of tissue environment during implantation.....	65
Table 3-7 Measured SAR over standard bands for health care telemetry systems.....	77
Table 3-8 Parameter values: Link budget calculation for a separation distance of 1m between the Tx and the Rx antennas.....	82
Table 4-1 Detailed dimension of the proposed Printed Folded Co-Planar Loop Antenna (Unit: Millimeters).....	105
Table 4-2 Size and performance comparison of the proposed Tx antenna.....	107
Table 4-3 Size and performance comparison of the proposed CP-PIFA Rx antenna.....	118
Table 4-4 Detailed dimension of the proposed CP-PIFA (Unit: Millimeters).....	118
Table 4-5 Radiation efficiencies of proposed antennas in this chapter	118
Table 4-6 Detailed dimension of the proposed CP-PIFA (Unit: Millimeters).....	126

Table 5-1 Human head tissue dielectric properties for different medical frequency bands	158
Table 5-2 SS-PIFAs Size and characteristic over different medical frequency bands	158
Table 5-3 Size comparison of proposed Single Slot PIFA	162
Table 5-4 Miniaturized SS-PIFAs size and characteristic over different medical frequency bands	163
Table 5-5 Losses in SS-PIFA and numerical phantom at 2.45 GHz (in watt).....	164
Table 5-6 Measured SAR over standard bands for health care telemetry systems..	172
Table 5-7 Detailed dimension of the proposed Folded Oval Loop antenna (Unit: Millimeters).....	176
Table 5-8 Size and performance comparison of the proposed Tx antenna	177
Table 5-9 Proposed Multi Band Single Slot PIFA features	181
Table 6-1 Dielectric properties of human body	199
Table 6-2 Optimized phase and amplitude distribution of the dual slot pifa array..	213
Table 6-3 Optimized phase and amplitude distribution of the dual slot pifa array..	231

LIST OF ARONYMS USED

Acronym	Expansion
ACMA	Australian Communications and Media Authority
ANSI	American National Standards Institute
A.R	Axial Ratio
AWGN	Additional White Gaussian Noise
BP	Blood Phantom
BS	Base Station
BSN	Body Sensor Node
CAD	Computer Aided Design
CP	Circular Polarization
CP-PIFA	Circular Polarization Planar Inverted-F Antenna
CSRR	Complementary Split Ring Resonator
CPW	Coplanar Waveguide
CSF	Cerebrospinal fluid
DGF	Dyadic Green's Function
DS-PIFA	Dual Slot Planar Inverted-F Antenna
EIRP	Equivalent Isotropically Radiated Power
EM	Electromagnetic
ERP	Effective radiated power
FCC	Federal Communications Commission
FDTD	Finite-difference Time Domain
FHBP	Full Human-Body Phantom
FPCB	Flexible Printed Circuit Board
GI	Gastrointestinal

HWD	Half-Wavelength Dipole
ICP	Intracranial Pressure Monitoring
IEEE	Institute of Electrical and Electronics Engineers
IMD	Implantable Medical Devices
LNA	Low Noise Amplifier
IOP	Intraocular pressure
ISM	Industrial, Scientific and Medical
LS	Least Squares
MBAN	Medical Body Area Network
MBSS-PIFA	Multi Band Single Slot Planar Inverted-F Antenna
MedRadio	Medical Device Radio communications Service
MICS	Medical Implant Communications Service
MLE	Maximum Likelihood Estimation
MRI	Magnetic Resonance Imaging
NI-LMH	Non-Invasive Local Microwave Hyperthermia
NFF	Near-field Focused
PL	Path Loss
PIFA	Planar Inverted-F Antenna
PDF	Probability Density Function
PDMS	Polydimethylsiloxane
PEC	Perfect Electric Conductors
RF	Radio Frequency
RFID	Radio Frequency Identification
RMSE	Root Mean Square Error
RSSI	Received Signal Strength Indicator
Rx	Receiver

SAR	Specific Absorption Rate
SLL	Side lobe level
SS-PIFA	Single Slot Planar Inverted-F Antenna
SRR	Split Ring Resonator
TAP	Transmitter Antenna Position
TBI	Traumatic Brain Injury
Tx	Transmitter
UWB	Ultra Wide Band
VHF	Very High Frequency
VNA	Vector Network Analyzer
WBAN	Wireless Body Area Networks
WCE	Wireless Capsule Endoscopy
WC-PLA	Wideband co-planar loop antenna
WFC-PLA	Wideband folded co-planar loop antenna
WFOLA	Wideband folded oval loop antenna
WMTS	Wireless Medical Telemetry Service
WOLA	Wideband oval loop antenna
WPC-PLA	Wideband printed co-planar loop antenna
WPOLA	Wideband printed oval loop antenna
WPT	Wireless Power Transfer

LIST OF SYMBOLS USED

Symbol	Quantity	Symbol	Quantity
α	Attenuation constant	n	Path loss exponent
β'	Phase constant	Pl	Path Loss
λ	Wavelength	ρ	Density
λ'	Wavelength in lossy mediums	$P_r(d)$	received power
ε	Dielectric constant	P_t	transmitted power
μ	Permeability	σ	Conductivity
G_t	Transmitter antenna gain	G_r	Receiver antenna gain
ϵ_r	Relative permittivity	G_c	Coding gain
S	Scattering parameter	G_d	Fixing deterioration
d	Tx-Rx separating distance	E_b	Energy per bit
$\tan \delta$	Loss tangent	R	Bit rate
l_m	Distance where the mutual coupling reduces significantly between the Tx and Rx antennas	X	Zero mean log-normally distributed random variable
L	Miscellaneous system losses	φ	Phase shift
N_0	Noise power spectral density	VF	Velocity factor
c_b	Blood specific heat	c	Specific heat
T_b	Blood Temperature	T	Temperature
η	Power transmission efficiency	F	Blood flow rate
L_{feed}	Antenna feeding loss	K	Thermal conductivity
C	Distributed capacitance	L	Distributed inductance
Z_0	Characteristic impedance	t	Time
λ_{eff}	Effective Wavelength	$E_{mn}(p)$	Array radiated field
I_0	Feeding current amplitude	F_0	Array focal point

ABSTRACT

Implantable medical devices (IMDs) introduced to monitor and transfer physiological information from inside the human body have superb potentials to provide major contributions to disease diagnosis, prevention and therapy. Moreover, minimally invasive biomedical devices helps to reduce the period of long-term hospitalization, so that enhancing the patients' quality of life. Understanding and developing biotelemetry devices, recording/transmitting data from inside the body to the external base station, requires a multi-disciplinary approach. Such a challenging task merges applied solutions, concepts and models from various fields, including biology, electronics, electromagnetism and package/system engineering. Among the device components, the transmitter antenna plays a key role.

Antenna design for biotelemetry applications is extremely challenging due to the effect of the surroundings on the radiator, the essential requirement to miniaturize the antenna structure size, reduced antenna efficiency and the robust effect of multipath losses. More specifically, in this thesis, I design and fabricate several antennas to be integrated in ingestible and implantable devices useful for remote monitoring as well as data biotelemetry. This work also focuses on arrays of body-worn antennas for both wireless endoscope base stations and cancer treatment nearfield microwave systems. Here, my aim is to reduce the physical size of the implantable antennas at specified operating standards frequency bands, while maintaining the antenna electromagnetic performance satisfactory. To achieve this, I introduce and use valuable miniaturization techniques for implantable patch antennas for biotelemetry applications.

Additionally, I design and fabricate compact microwave systems for cancer treatment using electromagnetic (EM) energy. Non-Invasive Local Microwave Hyperthermia (NI-LMH), which is my interest in this thesis, is a heat treatment serves to enhance the effectiveness of chemotherapy or radiation therapy and leads to gain remarkable results. The system may directly apply heat to a fairly small specific area such as tumors to destroy the local cancer cells. To achieve this, the heat effect is developed in the target by the transmission of EM energy, using array of antennas, which is adjusted in frequency, time and strength in order to work together to form a focus in the target. This places high demands on the precision of the system.

In this thesis, I present different planar antenna array for non-invasive microwave hyperthermia applications. The new Near Field Focused (NFF) arrays operates at ISM 2.45 GHz band and consists of 5 to 25 miniaturized dual slot PIFAs, depends on the array geometry arrangement. The arrays immersed inside a coupling bolus occupies a very small volume of space results in an easy fitting to contoured patient anatomy. These arrays, which are low profile and lightweight, have both superficial and deep focusing properties. The novel NFF body worn arrays are capable to focus on a single target with a high level of accuracy to concentrate the EM energy only on the target. I employ optimized dual slot PIFAs as array elements to reduce the size of the focusing area so that destroying very small tumors and avoid heating up the surrounding healthy cells.

I have also introduced size reduced NFF array to simplify the feeding network of the applicator and reduce the system cost, this array configuration satisfies system requirements and can focus on small targets precisely while keep the penetration depth high enough to heat up the deep seated targets.

In addition, the performance of both fabricated narrowband and wideband single antennas and array of antennas verified using experimental tissue mimicking phantoms. To validate the dielectric properties of experimental phantoms, over different frequency bands, dielectric probe kit employed, furthermore, optical fibre thermometers employed to confirm specific absorption rate (SAR) values for implanted radiators.

Index Terms – Antennas in lossy matter, Electrically small antennas, PIFA miniaturization method, SAR, Biocompatible insulation, Implantable and ingestible antennas, Rx and Tx antenna for wireless capsule endoscopy, Body centric wireless communications, compact antenna for head implants, Body Sensor Node (BSN), standards biotelemetry bands, array antenna for non-invasive local hyperthermia, Medical Device Radio communications Service (MedRadio), Medical Implanted Communication System (MICS).

This Page Intentionally Left Blank.